Edition Date: 2013-04-03 Revision Date: 2022-08-08

Revised by: Annette Roy, QA Coordinator

Product Name: HistoTainer™ II Catalogue No.: M961-90FW

This document replaces any previous version

1. **Product Description:**

> 90 ml Prefilled Specimen Container with White Lid: Disposable, non-sterile

Prefilled with 10% Neutral Buffered Formalin

2. Packaging:

Case: 4 internal trays of 24 units / 96 units per case

(Packaged in sturdy cardboard box with handles for easy carrying)

3. **Product Specifications:**

➤ Material:

• Container: Polypropylene

• White Lid: High Density Polyethylene

- Half prefilled with 10% Neutral Buffered Formalin (SDS available upon request)
- Wide lid ribbed cap for easy opening
- Closure with a unique integrated leak-resistant seal
- Cap configuration allows insertion of a Capinsert[™] (T345 Series) for color-coding

4. **Standards and Conformity:**

> ISO 2859-1: Sampling and inspection procedures Resin conforms to FDA 21 CFR 177.1520 > FDA: ➤ REACH (SVHC): Plastic is in conformity to REACH standards

5. **Quality Assurance:**

- > Clear, no presence of contamination in plastic
- Visual attributes
- > Dimension and volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg

Page 1 / 2 Issue Date: 2017-06-15

Document Template: FRM-106 Rev01 Technical Data Sheet

Edition Date: 2013-04-03 Revision Date: 2022-08-08

Revised by: Annette Roy, QA Coordinator

Traceability: 6.

Lot No. Composition : 8 or 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

7. **Storage Conditions:**

- > Store at room temperature in normal warehouse conditions. Store locked up. Store in a well-ventilated place. Keep container tightly closed
- > Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

8. **Recommended Use:**

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are Simport Scientific's copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Page 2 / 2 Issue Date: 2017-06-15

Document Template: FRM-106 Rev01 Technical Data Sheet