**Edition Date:** 2017-08-29 Revision Date: 2020-08-25

Revised by: Annette Roy, Regulatory

**Product Name: Transport Tube** Catalogue No.: T307-10A

This document replaces any previous version

# 1. Product Description:

➤ 10ml Q-Swab<sup>TM</sup> Transport Tube with External Thread: Disposable, Sterile, self-standing and non-

graduated

2. Packaging:

Case: 50 units per package / 500 units per case

### 3. Product Specifications:

Material:

• Tube: Polypropylene • Cap: Polyethylene • Washer Seal: Silicone

• Tipped Applicator: Polyester tip on polypropylene shaft

➤ Gamma radiation sterilized at a SAL of 10<sup>-3</sup>; specified dose between 6.5 kGy and 13.5 kGy

> Temperature range: -196°C to +121°C

### 4. Standards and Conformity:

> ISO 2859-1: Sampling and inspection procedures ➤ FDA: Resin conforms to FDA 21 CFR 177

> CONEG / RoHS: Plastics and colorants are in conformity with

CONEG / RoHS standards for heavy metals

> REACH (SVHC): Plastic is in conformity to REACH standards

> LATEX: Material is Latex Free ➤ BSE / TSE: Material is BSE / TSE Free

## 5. Quality Assurance:

- Clear, no presence of contamination in plastic
- Visual attributes
- > Volume measurements
- Closure verification
- ➤ Leak proof testing in vacuum at 71.3cm Hg
- Gas phase of Liquid Nitrogen resistance

Document Template: FRM-106 Rev01 Technical Data Sheet Page 1 / 2

Issue Date: 2017-06-15

**Edition Date:** 2017-08-29 **Revision Date:** 2020-08-25

**Revised by:** Annette Roy, Regulatory

## 6. Traceability:

**Lot No. Composition:** 8 or 9 digits

#### > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

# 7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- > Avoid temperature variations and humidity
- > Protect from any possible contamination
- > Protect from any damage to the packaging which could compromise the product sterility

#### 8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions
- Should be used only in the gas phase of Liquid Nitrogen

#### **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet
Page 2 / 2

Issue Date: 2017-06-15