Edition Date: 2005-10-08 Revision Date: 2017-05-06

Revised by: Annette Roy, Regulatory

Product Name: Micrewtube® Catalogue No.: T332-4SPR

This document replaces any previous version

1. **Product Description:**

> > 1.5ml Micrewtube®: Sterile, Printed, self-standing tube with Silicone washer seal screw

> > cap and attachment loop. Caps are pre-attached and screwed on.

2. Packaging:

> Case: 10 bags of 50 for a total of 500 units per case

3. **Product Specifications:**

- ➤ Material:
 - Polypropylene tube
 - Polypropylene cap
 - Silicone washer seal
- Certified RNase, DNase, Pyrogen and DNA Free
- > Gamma radiation sterilized at a SAL of 10⁻³; specified dose between 6.5 kGy and 13.5 kGy
- ➤ Temperature range: -196°C to +121°C.
- > Autoclavable at 121°C, for up to 30 minutes
- > Tubes have graduations and white marking area
- > Centrifuged up to 17 000g

4. **Standards and Conformity:**

> ISO 2859-1: Sampling and inspection procedures > FDA: Resin conforms to FDA 21 CFR 177.1520

➤ USP: Resin conforms to USP Class VI

> CONEG: Plastics and colorants are in conformity with

CONEG standards for heavy metals

Plastic is in conformity to REACH standards > REACH (SVHC):

> LATEX: Material is Latex Free ➤ BSE / TSE: Material is BSE / TSE Free

5. **Quality Assurance:**

- Clear, no presence of contamination in plastic
- Visual attributes
- Volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg
- Gas phase of Liquid Nitrogen resistance

Document Template: FRM-106 Rev01 Technical Data Sheet Page 1 / 2

Issue Date: 2017-06-15

Edition Date: 2005-10-08 **Revision Date:** 2017-05-06

Revised by: Annette Roy, Regulatory

6. Traceability:

Lot No. Composition: 8 or 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

7. Storage Conditions:

- Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions
- > Should be used only in the gas phase of Liquid Nitrogen

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet
Page 2 / 2

Issue Date: 2017-06-15