**Edition Date:** 2005-10-08 **Revision Date:** 2021-10-26

**Revised by:** Annette Roy, Regulatory

Product Name: Micrewtube® Catalogue No.: T338-7

This document replaces any previous version

# 1. Product Description:

➤ 2 ml Micrewtube®: Non-Sterile, conical bottom tube with lip seal

screw cap. Non-assembled separate caps and

non-printed tubes

2. Packaging:

> Case: bag of 1000 tubes and bag of 1000 caps / 1000 units per case

# 3. Product Specifications:

- Material:
  - Polypropylene tube
  - Polyethylene cap
- Certified RNase, DNase, Pyrogen and DNA Free
- > Temperature range: -196°C to +100°C
- Tubes have white marking areas
- Centrifuged up to 20 000 g

# 4. Standards and Conformity:

➤ ISO 2859-1: Sampling and inspection procedures➤ FDA: Resin conforms to FDA 21 CFR 177.1520

> USP: Resin conforms to USP Class VI

> CONEG / RoHS: Plastics and colorants are in conformity with

CONEG / RoHS standards for heavy metals

REACH (SVHC): Plastic is in conformity to REACH standards

LATEX: Material is Latex FreeBSE / TSE: Material is BSE / TSE Free

# 5. Quality Assurance:

- Clear, no presence of contamination in plastic
- Visual attributes
- Volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg
- Gas phase of Liquid Nitrogen resistance

Document Template: FRM-106 Rev00 Technical Data Sheet

Issue Date: 2016-04-26

Page 1 / 2

**Edition Date:** 2005-10-08 **Revision Date:** 2021-10-26

**Revised by:** Annette Roy, Regulatory

### 6. Traceability:

**Lot No. Composition:** 8 to 9 digits

### > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

# 7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- Protect from any damage to the packaging

#### 8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- For use in automated equipment, follow the equipment manufacturer's instructions
- Should be used only in the gas phase of Liquid Nitrogen

#### **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

 ${\tt Document\ Template:\ FRM-106\ Rev00\ Technical\ Data\ Sheet}$ 

Issue Date: 2016-04-26