

Technical Data Sheet

Edition Date: 2005-10-08 **Revision Date:** 2020-08-26

Revised by: Annette Roy, Regulatory

Product Name: Micrewtube® Catalogue No.: T339-2S

This document replaces any previous version

1. Product Description:

O.5ml Micrewtube®: Sterile, self-standing with lip-seal and flat-top cap.

Caps are screwed on.

2. Packaging:

Case: 10 packages of 50 units / 500 units per case

3. Product Specifications:

- Material:
 - Polypropylene tube
 - Polyethylene cap
- Certified RNase, DNase, Pyrogen and DNA Free
- ➤ Gamma radiation sterilized at a SAL of 10⁻³; specified dose between 6.5 kGy and 13.5 kGy
- > Temperature range: -196°C to +100°C
- > Tubes have no graduations and have no writing areas
- Centrifuged up to 17 000 g

4. Standards and Conformity:

ISO 2859-1: Sampling and inspection procedures
FDA: Resin conforms to FDA 21 CFR 177.1520

> USP: Resin conforms to USP Class VI

> CONEG / RoHS: Plastics and colorants are in conformity with

CONEG / RoHS standards for heavy metals

REACH (SVHC): Plastic is in conformity to REACH standards

LATEX: Material is Latex FreeBSE / TSE: Material is BSE / TSE Free

5. Quality Assurance:

- Clear, no presence of contamination in plastic
- Visual attributes
- Volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg
- Gas phase of Liquid Nitrogen resistance

Document Template: FRM-106 Rev00 Technical Data Sheet

Issue Date: 2016-04-26

Page 1 / 2



Technical Data Sheet

Edition Date: 2005-10-08 **Revision Date:** 2020-08-26

Revised by: Annette Roy, Regulatory

6. Traceability:

Lot No. Composition: 8 to 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

7. Storage Conditions:

- Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging, which could compromise the product sterility

8. Recommended Use:

- Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- For use in automated equipment, follow the equipment manufacturer's instructions
- Should be used only in the gas phase of Liquid Nitrogen

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

 ${\tt Document\ Template:\ FRM-106\ Rev00\ Technical\ Data\ Sheet}$

Issue Date: 2016-04-26