

Edition Date: Revision Date : Revised by: 2004-02-05 2017-11-03 Annette Roy, QA Coordinator

# Product Name: Micrewtube<sup>®</sup> Catalogue No.: T341-6TLST

#### This document replaces any previous version

**1. Product Description:** 

2ml Micrewtube®:

Non sterile, Low adhesion, self-standing, disposable tube. Cap sold separately

- > Compatible with:
  - T340 Cap Series
- 2. Packaging:
  - Case: 1 000 units per bag

## 3. Product Specifications:

- > Material:
  - Polypropylene with Low Adhesion properties
- Certified RNase, DNase, Pyrogen and DNA Free
- Temperature range: -196°C to +121°C
- Autoclavable at 121°C, for up to 30 minutes
- Centrifuge resistant at up to 17 000 g

### 4. Standards and Conformity:

- ISO 2859-1: Sampling and inspection procedures
- FDA: Resin conforms to FDA 21 CFR 177.1520
- USP: Resin conforms to USP Class VI
- > CONEG / RoHS: Plastics and colorants are in conformity with
- CONEG / RoHS standards for heavy metals
- **REACH (SVHC):** Plastic is in conformity to REACH standards

Material is BSE / TSE Free

- LATEX: Material is Latex Free
- BSE / TSE:



Edition Date: Revision Date : Revised by: 2004-02-05 2017-11-03 Annette Roy, QA Coordinator

## 5. Quality Assurance:

- Clear, no presence of contamination in plastic
- Visual attributes
- Volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg
- Gas phase of Liquid Nitrogen resistance

## 6. Traceability:

**Lot No. Composition:** 8 or 9 digits

### > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

## 7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

### 8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions
- Should be used only in the gas phase of Liquid Nitrogen

### **OWNERSHIP OF MATERIALS**:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.