

Edition Date: Revision Date : Revised by: 2014-06-18 2020-09-08 Annette Roy, Regulatory

Product Name: Micrewtube[®] Catalogue No.: T361-7TPR

This document replaces any previous version

- **1. Product Description:**
- \triangleright

2 ml Printed Micrewtube®:

Non-sterile, Disposable conical bottom tube with molded ridges, printed marking area and graduations. Caps sold separately.

- > Compatible with:
 - T340 Cap Series
 - T360 OneHand[™] Rack

2. Packaging:

Case: 1 000 units per bag

3. Product Specifications:

- ➢ Material: Polypropylene
- > These tubes have molded ridges matching serrations on rack wells
- Certified RNase, DNase, Pyrogen and DNA Free
- Temperature range: -196°C to +121°C
- Autoclavable at 121°C, for up to 30 minutes
- Centrifuge resistant at up to 17 000 g

4. Standards and Conformity:

\triangleright	ISO 2859-1:	Sampling and inspection procedures
------------------	-------------	------------------------------------

FDA: Resin conforms to FDA 21 CFR 177.1520
USP: Resin conforms to USP Class VI
CONEG / RoHS: Plastics and colorants are in conformity with CONEG / RoHS standards for heavy metals
REACH (SVHC): Plastic is in conformity to REACH standards
LATEX: Material is Latex Free
BSE / TSE: Material is BSE / TSE Free



Edition Date: Revision Date : Revised by: 2014-06-18 2020-09-08 Annette Roy, Regulatory

5. Quality Assurance:

- Clear, no presence of contamination in plastic
- Visual attributes
- Volume measurements
- Closure verification
- > Leak proof testing in vacuum at 71.3cm Hg
- ➢ Gas phase of Liquid Nitrogen resistance

6. Traceability:

Lot No. Composition: 8 or 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions
- Should be used only in the gas phase of Liquid Nitrogen

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.