**Edition Date:** 2018-07-27 **Revision Date:** NONE

Revised by: Annette Roy, QA Coordinator

Page 1 / 2

Product Name: MacrewTube™ Catalogue No.: T365-5T series

This document replaces any previous version

1. **Product Description:** 

> 5ml MacrewTube: Non Sterile, Threaded conical bottom tube. Cap sold

> > Separately

> Available colors:

Natural and Amber

2. Packaging:

> Case: 2 packages of 100 units / 200 units per case

#### **Product Specifications:** 3.

- Material:
  - Polypropylene
- Cap has a locking system to prevent opening during boiling applications
- Cap can be pierced easily with a syringe needle
- Certified RNase, DNase, Pyrogen and DNA Free
- Temperature range: -196°C to +121°C.
- Autoclavable at 121°C, for up to 30 minutes
- Molded graduation markings incremented at every 0.5ml and up to 5ml
- Etched surface on side of tube for sample identification
- Centrifuge resistant at up to 25 000g

## **Standards and Conformity:**

> ISO 2859-1: Sampling and inspection procedures > FDA: Resin conforms to FDA 21 CFR 177.1520 > CONEG / RoHS: Plastics and colorants are in conformity with CONEG / RoHS standards for heavy metals

Plastic is in conformity to REACH standards

> LATEX: Material is Latex Free **▶** BSE / TSE: Material is BSE / TSE Free

#### 5. **Quality Assurance:**

> REACH (SVHC):

- Clear, no presence of contamination in plastic
- Visual attributes
- Closure verification

Document Template: FRM-106 Rev00 Technical Data Sheet Issue Date: 2016-04-26

Edition Date: 2018-07-27 Revision Date: NONE

**Revised by:** Annette Roy, QA Coordinator

# 6. Traceability:

Lot No. Composition: 8 to 9 digits

## > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

# 7. Storage Conditions:

- Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

### 8. Recommended Use:

- Verify proper cap closure when using biohazard material and / or chemical reagents;
- > Follow chemical resistance chart recommendations;

### **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev00 Technical Data Sheet

Issue Date: 2016-04-26