**Edition Date:** 2012-07-18 **Revision Date:** 2020-08-26

**Revised by:** Annette Roy, Regulatory

Product Name: Cultubes<sup>™</sup> Catalogue No.: T405-2A

This document replaces any previous version

### 1. Product Description:

> 5ml Culture Tube with Cap: Disposable, sterile, 12 x 75mm culture tube with graduations

and white marking area

2. Packaging:

Case: Packaged in 20 trays of 25 / 500 units per case

# 3. Product Specifications:

Material:

Tube: PolypropyleneCap: PolyethyleneDual-position snap cap:

- Sealed position for anaerobic cultures
- Loose position for aerobic work
- Certified Pyrogen Free
- ➤ Gamma radiation sterilized at a SAL of 10<sup>-3</sup>; specified dose between 6.5 kGy and 13.5 kGy
- > Temperature range: -90°C to +100°C
- ➤ Not autoclavable
- Centrifuge resistant at up to 3 000 g

### 4. Standards and Conformity:

➤ ISO 2859-1: Sampling and inspection procedures
 ➤ FDA: Resins conform to FDA 21 CFR 177.1520
 ➤ CONEG / RoHS: Plastics and colorants are in conformity with CONEG / RoHS standards for heavy metals

> REACH (SVHC): Plastics are in conformity to REACH standards

LATEX: Materials are Latex FreeBSE / TSE: Materials are BSE / TSE Free

### 5. Quality Assurance:

- > Clear, no presence of contamination in plastic
- Visual attributes
- > Dimension and volume measurements

Closure verification

Document Template: FRM-106 Rev01 Technical Data Sheet

Issue Date: 2017-06-15

Page 1 / 2

**Edition Date:** 2012-07-18 Revision Date: 2020-08-26

Revised by: Annette Roy, Regulatory

#### 6. Traceability:

> Lot No. Composition: 8 or 9 digits

## > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

### **Storage Conditions:** 7.

- > Store at room temperature in normal warehouse conditions
- > Avoid temperature variations and humidity
- > Protect from any possible contamination
- Protect from any damage to the packaging which could compromise the product sterility

#### 8. **Recommended Use:**

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions

### **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are Simport Scientific's copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet Page 2 / 2

Issue Date: 2017-06-15