Edition Date: 2006-03-28 Revision Date: 2020-07-17

Revised by: Annette Roy, Regulatory

Product Name: Cultubes™ Catalogue No.: T426-2A

This document replaces any previous version

1. **Product Description:**

> 14 ml Culture Tube with Cap: Disposable, sterile 17 x 95 mm culture tube, non-printed. Assembled with Cap.

2. Packaging:

> Case: 20 bags of 25 units / 500 units per case

3. **Product Specifications:**

- Material:
 - Tube: Polypropylene
 - Cap: High Density Polyethylene
- > Dual-position snap cap:
 - Sealed position for anaerobic cultures
 - Loose position for aerobic work
- Certified Pyrogen Free
- > Gamma radiation sterilized at a SAL of 10⁻³; specified dose between 6.5 kGy and 13.5 kGy
- > Temperature range tube: -80°C to +80°C
- ➤ Not autoclavable
- Centrifuge resistant at up to 3 000 g

4. **Standards and Conformity:**

> ISO 2859-1: Sampling and inspection procedures > FDA: Resins conform to FDA 21 CFR 177.1520 > CONEG / RoHS: Plastics and colorants are in conformity with CONEG / RoHS standards for heavy metals

Plastic is in conformity to REACH standards > REACH (SVHC):

> LATEX: Material is Latex Free ➤ BSE / TSE: Material is BSE / TSE Free

5. **Quality Assurance:**

- > Clear, no presence of contamination in plastic
- Visual attributes
- > Dimension and volume measurements
- Closure verification

Page 1 / 2 Issue Date: 2017-06-15

Document Template: FRM-106 Rev01 Technical Data Sheet

Edition Date: 2006-03-28 **Revision Date:** 2020-07-17

Revised by: Annette Roy, Regulatory

6. Traceability:

Lot No. Composition: 8 or 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- > Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging which could compromise the product sterility

8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet
Page 2 / 2

Issue Date: 2017-06-15