

**Edition Date:** 2008-07-22

Edited by: Annette Roy, QA Coordinator

**Revision date:** 2022-03-29

Revised by Christine Laforge RT, Technical Representative

# **Technical Data Sheet**

**Product Name: Sample Tube** Catalogue No.: T501-5ATPR

This document replaces any previous version

### 1. **Product Description:**

> 5ml Sample Tube with External Thread: Non-sterile, disposable, Self-standing, printed, external

threaded design. Caps are sold separately.

Compatible with:

• T501 & T502 Screw Cap Series

2. Packaging:

> Case: 1 000 units per package

#### 3. **Product Specifications:**

➤ Material:

Polypropylene

> Temperature range: -196°C to +121°C

> Tubes can be autoclaved in upright position with caps loosened at +121°C, for up to 30 minutes

➤ Gamma radiation resistant

#### 4. **Standards and Conformity:**

> ISO 2859-1: Sampling and inspection procedures ➤ FDA: Resin conforms to FDA 21 CFR 177.1520

> USP: Resin conforms to USP Class VI

> CONEG / RoHS: Plastics and colorants are in conformity with

CONEG / RoHS standards for heavy metals

> REACH (SVHC): Plastic is in conformity to REACH standards

> LATEX: Material is Latex Free

➤ BSE / TSE: Material is BSE / TSE Free

#### 5. **Quality Assurance:**

- Clear, no presence of contamination in plastic
- Visual attributes
- Dimension and volume measurements
- Closure verification
- ➤ Leak proof testing in vacuum at 71.3cm Hg
- Gas phase of Liquid Nitrogen resistance

Document Template: FRM-106 Rev01 Technical Data Sheet Page 1 / 2

Issue Date: 2017-06-15



**Edition Date:** 2008-07-22

Edited by: Annette Roy, QA Coordinator

**Revision date:** 2022-03-29

Revised by Christine Laforge RT, Technical Representative

# **Technical Data Sheet**

### 6. Traceability:

> Lot No. Composition: 8 or 9 digits

## > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

### 7. **Storage Conditions:**

- > Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

#### 8. **Recommended Use:**

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions
- > Should be used only in the gas phase of Liquid Nitrogen

## **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are Simport Scientific's copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet Page 2 / 2

Issue Date: 2017-06-15