

# Scientific inc. Technical Data Sheet

**Edition Date:** 2022-02-16 **Revision Date:** 2022-08-08

Christine Laforge RT, Revised by: Représentante technique

Product Name: HistoTainer™ II Catalogue No.: M961-20FW

This document replaces any previous version

### 1. **Product Description:**

**20ml Prefilled Specimen Container with White Lid:** Disposable, non-sterile

Prefilled with 10% Neutral Buffered Formalin

### 2. Packaging:

Case: 4 internal trays of 24 units / 96 units per case

(Packaged in sturdy cardboard box with handles for easy carrying)

### 3. **Product Specifications:**

Material:

• Container: Polypropylene

• White Lid: High Density Polyethylene

- Half prefilled with 10% Neutral Buffered Formalin (SDS available upon request)
- Wide lid ribbed cap for easy opening
- Closure with a unique integrated leak-resistant seal
- ➤ Cap configuration allows insertion of a Capinsert<sup>TM</sup> (T345 Series) for color-coding

#### 4. **Standards and Conformity:**

Sampling and inspection procedures > ISO 2859-1: > FDA: Resin conforms to FDA 21 CFR 177.1520 > REACH (SVHC): Plastic is in conformity to REACH standards

#### 5. **Quality Assurance:**

- > Clear, no presence of contamination in plastic
- Visual attributes
- > Dimension and volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg

Page 1 / 2 Issue Date: 2017-06-15

Document Template: FRM-106 Rev01 Technical Data Sheet



# Scientific inc Technical Data Sheet

**Edition Date:** 2022-02-16 **Revision Date:** 2022-08-08

**Revised by:** Christine Laforge RT,
Représentante technique

# 6. Traceability:

Lot No. Composition: 8 or 9 digits

### > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

## 7. Storage Conditions:

- Store at room temperature in normal warehouse conditions. Store locked up. Store in a well-ventilated place. Keep container tightly closed
- > Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

## 8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions

### **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet
Page 2 / 2

Issue Date: 2017-06-15