Edition Date: 2016-09-15 Revision Date: 2022-08-08

Revised by: Annette Roy, QA Coordinator

Product Name: Urisafe® Catalogue No.: **B350**

This document replaces any previous version

1. **Product Description:**

> 3 Liters, 24-Hr Urine Container: Disposable, Non-sterile assembled with screw cap with washer

liner.

2. Packaging:

> Case: 40 units per case

3. **Product Specifications:**

Material:

• Bottle: Polyethylene • Cap: Polypropylene • Valve: Polyethylene Gamma radiation resistant

> Temperature range: -80°C to +100°C.

4. **Standards and Conformity:**

> ISO 2859-1: Sampling and inspection procedures Resin conforms to FDA 21 CFR 177.1520 > FDA: CONEG / RoHS: Plastics and colorants are in conformity with

CONEG /RoHS standards for heavy metals

Material is Latex free > LATEX:

5. **Quality Assurance:**

- No presence of contamination in plastic;
- Visual attributes:
- Closure and valve verification.

Page 1 / 2 Issue Date: 2017-06-15

Document Template: FRM-106 Rev01 Technical Data Sheet

Edition Date: 2016-09-15 Revision Date: 2022-08-08

Revised by: Annette Roy, QA Coordinator

6. Traceability:

> Lot No. Composition: 8 or 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case;
- 3. On the inner bag.

7. **Storage Conditions:**

- > Store at room temperature in normal warehouse conditions
- Protect from any possible contamination
- Protect from any damage to the packaging

8. **Recommended Use:**

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are Simport Scientific's copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet Page 2 / 2

Issue Date: 2017-06-15