

2016-11-03 2022-04-26 Christine Laforge RT, Technical Representative

| Product Name: Tamper Evident Transport Tube | Catalogue No.: T552-30ATTP |
|--|----------------------------|
| This document replaces any previous version | |
| 1 Product Description: | |

1. Product Description:

30 ml Tamper Evident Transport Tube: Disposable, non-sterile, self-standing and non-graduated. Caps sold separately.

> Compatible with:

- T552WTP Tamper Evident Cap with Lip Seal
- T552W Lip Seal Cap

2. Packaging:

Case:1 000 units per pack

3. Product Specifications:

- > Material: Polypropylene
- Temperature range: -90°C to +121°C
- > Autoclavable at 121°C, for up to 30 minutes
- > Tube has a tamper evident collar
- Gamma radiation resistant

4. Standards and Conformity:

- ISO 2859-1:
 - -1: Sampling and inspection procedures Resin conforms to FDA 21 CFR 177.1520
- FDA: Resin conforms to FDA 21 CFR 177.1520
 CONEG / RoHS: Plastics and colorants are in conformity with CONEG / RoHS standards for heavy metals
 REACH (SVHC): Plastic is in conformity to REACH standards
- LATEX: Material is Latex Free
- **BSE / TSE:** Material is BSE / TSE Free

5. Quality Assurance:

- > Clear, no presence of contamination in plastic
- Visual attributes
- Dimensional and volume measurements
- Closure verification
- ➢ Gas phase of Liquid Nitrogen resistance
- Leak proof testing in vacuum at 71.3cm Hg (95 kPa tested)



Edition Date: Revision Date : Revised by: 2016-11-03 2022-04-26 Christine Laforge RT, Technical Representative

- 6. Traceability:
 - Lot No. Composition: 8 or 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- > Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.