

## Scientific inc Technical Data Sheet

**Edition Date:** 2004-03-15 Revision Date: 2022-10-11

Revised by: Christine Laforge RT, Technical Representative

**Product Name: Dropette®** Catalogue No.: P200-14

This document replaces any previous version

1. **Product Description:** 

> > Transfer Pipet: Non-sterile, disposable pipet

2. Packaging:

> Case: 10 inner boxes of 500 units / 5 000 units per case

3. **Product Specifications:** 

Material:

• Low Density Polyethylene

> Graduated increments from 0.1 ml to 0.3 ml

Bulb draw: 0.8ml > Total capacity: 1.5ml > Length: 11.4 cm

> Temperature range : -80°C to +80°C

4. **Standards and Conformity:** 

> > ISO 2859-1: Sampling and inspection procedures ➤ FDA: Resin conforms to FDA 21 CFR 177.1520

➤ Latex: Material is Latex free

5. **Quality Assurance:** 

- Clear, no presence of contamination in plastic
- Visual attributes
- > Dimension and volume measurements

6. Traceability:

> **➢** No. Composition: 8 or 9 digits

> The lot number can be found in one or all of these locations:

1. On exterior case label

2. On label inserted inside the master case

3. On the inner bag

Document Template: FRM-106 Rev01 Technical Data Sheet Page 1 / 2 Issue Date: 2017-06-15

**Edition Date:** 2004-03-15 **Revision Date:** 2022-10-11

Revised by: Christine Laforge RT, Technical Representative

## **Storage Conditions:** 7.

- > Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any damage to the packaging

## 8. Recommended Use:

- > Be cautious when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations

## **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are Simport Scientific's copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet Page 2 / 2

Issue Date: 2017-06-15