

# Scientific inc Technical Data Sheet

**Edition Date:** 2022-11-10 **Revision Date:** None

Revised by: Christine Laforge RT, Technical Representative

Product Name: Histosette® II Catalogue No.: M483- BA Series

## 1. **Product Description:**

Biopsy Cassette Base: Non-Sterile, disposable, compatible with lid M483, E-Z Load cassettes

# > Available Colors:

- White, Pink, Green, Yellow, Blue, Peach, Tan, Gray, Lilac, Orange and Aqua
- Fluorescent Pink, Green and Yellow also available, minimum order quantity required

#### 2. Packaging:

Case: 20 stacks of 50 units / 2 packages per case: total 2 000 units

#### 3. **Product Specifications:**

- Material:
  - Acetal
  - Cassettes base are attached with a zip tie
  - Cassettes base are compatible with Shur/Mark® Cassette Labeling Instruments

#### 4. **Standards and Conformity:**

> ISO 2859-1: Sampling and inspection procedures > FDA: Resin conforms to FDA 21 CFR 177.2470 CONEG / RoHS: Plastics and colorants are in conformity with CONEG / RoHS standards for heavy metals

Plastic is in conformity to REACH standards > REACH (SVHC):

> LATEX: Material is Latex Free ➤ CE: Product is CE marked

## 5. **Quality Assurance:**

- > Clear, no presence of contamination in plastic
- Visual attributes
- ➤ Lid closure verification

Page 1 / 2 Issue Date: 2017-06-15

Document Template: FRM-106 Rev01 Technical Data Sheet

**Edition Date:** 2022-11-10 **Revision Date:** None

Revised by: Christine Laforge RT, Technical Representative

#### 6. **Traceability:**

**Lot No. Composition:** 8 or 9 digits

# > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

## 7. **Storage Conditions:**

- Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

#### 8. **Recommended Use:**

- > Verify proper lid closure of the cassette before using
- > Avoid picric acid, prolonged use in this acid will cause degradation
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions

## **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are Simport Scientific's copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Page 2 / 2 Issue Date: 2017-06-15

Document Template: FRM-106 Rev01 Technical Data Sheet