

Edition Date:2004Revision Date :2022Revised by:Ann

2004-02-09 2022-08-08 Annette Roy, QA Coordinator

# Product Name: Sample Tube Catalogue No.: C572-2

This document replaces any previous version

### **1.** Product Description:

> 30ml Sample Tube with Cap: Disposable, non-sterile and conical bottom tube with self-standing base

#### 2. Packaging:

Case: 5 packs of 100 units / 500 units per case

#### **3. Product Specifications:**

- > Material:
  - Tube: Polypropylene
  - Cap: Polyethylene
  - Cap Color: white
- Gamma radiation resistant
- ➤ Temperature range: -90°C to +100°C
- > Tubes have molded-in graduations from 10ml to 30ml with 5ml increments
- > Cap configuration allows insertion of a Capinsert<sup>™</sup> (T345 Series) for color-coding

## 4. Standards and Conformity:

- ISO 2859-1: Sampling and inspection procedures
- FDA: Resins conform to FDA 21 CFR 177.1520
- USP: Resins conform to USP Class VI
- > CONEG / RoHS: Plastics and colorants are in conformity with
  - CONEG / RoHS standards for heavy metals
- > **REACH (SVHC):** Plastics are in conformity to REACH standards
- LATEX: Materials are Latex Free
- > BSE / TSE:Materials are BSE / TSE Free

# 5. Quality Assurance:

- Clear, no presence of contamination in plastic
- Visual attributes
- Volume measurements
- Closure verification
- Gas phase of Liquid Nitrogen resistance



Edition Date: Revision Date : Revised by: 2004-02-09 2022-08-08 Annette Roy, QA Coordinator

# 6. Traceability:

- **Lot No. Composition:** 8 or 9 digits
- > The lot number can be found in one or all of these locations:
  - 1. On exterior case label
  - 2. On label inserted inside the master case
  - 3. On the inner bag

# 7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- Protect from any possible contamination
- Protect from any damage to the packaging

# 8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > Follow the equipment manufacturer's instructions
- > Should be used only in the gas phase of Liquid Nitrogen

## **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.