Edition Date: 2017-09-13 **Revision Date:** 2021-04-28

Revised by: Annette Roy, QA Coordinator

Product Name: BioTube™ Rack Catalogue No.: T101-1

This document replaces any previous version

1. Product Description:

> Rack with 1.2 ml tubes: Disposable, non-sterile, rack pre-filled with 96 plain individual tubes.

Caps sold separately.

> Rack Made of 3 Components:

- A white base with alphanumerical grid
- A translucent cover with alphanumerical grid
- 96 individual tubes

> Compatible with:

T100-28 Individual Plug Caps

2. Packaging:

Case: 2 packages of 5 units / 10 units per case

3. Product Specifications:

- Material:
 - Polypropylene
- Certified RNase, DNase, Pyrogen and DNA Free
- > Temperature range: -80°C to +121°C
- ➤ Autoclavable at +121°C, for up to 30 minutes
- > Cover and rack have easy to read alphanumerical dots
- ➤ Tubes total volume of 1.2 ml will hold 1.1 ml with cap
- ➤ Note: These racks size have a standard microtiter configuration

4. Standards and Conformity:

➤ ISO 2859-1: Sampling and inspection procedures➤ FDA: Resin conforms to FDA 21 CFR 177.1520

CONEG / RoHS:
Plastics and colorants are in conformity with CONEG

/ RoHS standards for heavy metals

➤ **REACH (SVHC):** Plastic is in conformity to REACH standards

LATEX: Material is Latex FreeBSE / TSE: Material is BSE / TSE Free

5. Quality Assurance:

Issue Date: 2017-06-15

Document Template: FRM-106 Rev01 Technical Data Sheet

Edition Date: 2017-09-13 **Revision Date:** 2021-04-28

Revised by: Annette Roy, QA Coordinator

- Clear, no presence of contamination in plastic
- Visual attributes
- Gas phase of Liquid Nitrogen resistance

6. Traceability:

Lot No. Composition: 8 or 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

7. Storage Conditions:

- Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

8. Recommended Use:

- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions
- > Should be used only in the gas phase of Liquid Nitrogen

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Issue Date: 2017-06-15

Document Template: FRM-106 Rev01 Technical Data Sheet