**Edition Date:** 2017-09-13 Revision Date: 2021-04-28

Revised by: Annette Roy, QA Coordinator

Product Name: BioTube™ Rack Catalogue No.: T101-3

This document replaces any previous version

#### 1. **Product Description:**

> Rack with 1.2 ml tubes: Disposable, non-sterile, rack pre-filled with 12 strips of 8 tubes. Caps

sold separately.

## > Rack Made of 3 Components:

- A white base with alphanumerical grid
- A translucent cover with alphanumerical grid
- 12 strips of 8 tubes

## Compatible with:

- T100-30 Strip of 8 Plug Caps
- T100-28 Individual Plug Caps

#### 2. Packaging:

Case: 2 packages of 5 units / 10 units per case

#### 3. **Product Specifications:**

- Material:
  - Polypropylene
- Certified RNase, DNase, Pyrogen and DNA Free
- > Temperature range: -80°C to +121°C
- > Autoclavable at +121°C, for up to 30 minutes
- Cover and rack have easy to read alphanumerical dots
- > Tubes total volume of 1.2 ml will hold 1.1 ml with cap
- ➤ Note: These racks size have a standard microtiter configuration

#### 4. **Standards and Conformity:**

> ISO 2859-1: Sampling and inspection procedures Resin conforms to FDA 21 CFR 177.1520 > FDA:

> CONEG / RoHS: Plastics and colorants are in conformity with CONEG

/ RoHS standards for heavy metals

> REACH (SVHC): Plastic is in conformity to REACH standards

> LATEX: Material is Latex Free ➤ BSE / TSE: Material is BSE / TSE Free

Page 1 / 2 Issue Date: 2017-06-15

**Edition Date:** 2017-09-13 **Revision Date:** 2021-04-28

Revised by: Annette Roy, QA Coordinator

## 5. Quality Assurance:

- Clear, no presence of contamination in plastic
- Visual attributes
- Gas phase of Liquid Nitrogen resistance

## 6. Traceability:

Lot No. Composition: 8 or 9 digits

### > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

# 7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- > Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

### 8. Recommended Use:

- > Follow chemical resistance chart recommendations
- For use in automated equipment, follow the equipment manufacturer's instructions
- Should be used only in the gas phase of Liquid Nitrogen

### **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet
Page 2 / 2

Issue Date: 2017-06-15