

## **Technical Data Sheet**

**Edition Date:** 2019-10-15 **Revision Date:** 2021-04-13

**Revised by:** Annette Roy, QA Coordinator

Product Name: Amplate<sup>™</sup> Catalogue No.: T324-384SK series

This document replaces any previous version

#### 1. Product Description:

> 384 wells Thin Wall PCR Plate: 40 ul volume capacity, Non-sterile

Available colors:

T324-384SKW : whiteT324-384SKK : Black

## **Compatible with:**

SecureSeal<sup>™</sup> Sealing Films and Foils (T329 Series);

#### 2. Packaging:

Case: 10 bags of 10 units packaged in resealable bags / 100 units per case

#### 3. Product Specifications:

- Material:
  - Polypropylene
- Certified RNase, DNase, Pyrogen and DNA Free
- Compatible with multi-channel pipettor
- ➤ Temperature range: -80°C to +121°C
- ➤ Autoclavable at +121°C for up to 30 minutes

#### 4. Standards and Conformity:

ISO 2859-1: Sampling and inspection procedures
FDA: Resin conforms to FDA 21 CFR 177.1520
CONEG / RoHS: Plastics and colorants are in conformity with CONEG / RoHS standards for heavy metals

Plastic is in conformity to REACH standards

> REACH (SVHC): Plastic is in conformity to REACH st

BSE / TSE: Material is BSE / TSE FreeLATEX: Material is Latex Free

Document Template: FRM-106 Rev00 Technical Data Sheet

Issue Date: 2016-04-26



# **Technical Data Sheet**

**Edition Date:** 2019-10-15 **Revision Date:** 2021-04-13

**Revised by:** Annette Roy, QA Coordinator

## 5. Quality Assurance:

- Clear, no presence of contamination in plastic
- Visual attributes

#### 6. Traceability:

**Lot No. Composition:** 8 to 9 digits

#### > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

#### 7. Storage Conditions:

- Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging, which could compromise the product integrity

#### 8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations

#### **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev00 Technical Data Sheet

Issue Date: 2016-04-26