Edition Date: 2014-05-21 **Revision Date:** 2020-08-04

Revised by: Annette Roy, QA Coordinator

Page 1 / 2

Product Name: Microcentrifuge Tube

Catalogue No.: T330-8

This document replaces any previous version

1. **Product Description:**

1.5ml Microcentrifuge Tube with pick-up tab: Non sterile, conical bottom, with attached cap

> Available colors:

 Natural, *Blue, *Green and *Yellow *colors available on request with minimum quantities applicable

2. Packaging:

Case: 10 packages of 500 units / 5 000 units per case

(packaged in tamperproof resealable zip-lock bags)

3. **Product Specifications:**

- Material:
 - Polypropylene
- Cap is snug fitting to prevent opening during boiling applications
- Certified RNase, DNase, Pyrogen and DNA Free
- Temperature range: -80°C to +121°C.
- Autoclavable at 121°C, for up to 30 minutes
- ➤ Molded graduation markings at 0.5ml, 1.0ml, and 1.5ml
- Etched surface on side of tube for sample identification
- Centrifuge resistant at up to 20 000g

Standards and Conformity:

> ISO 2859-1: Sampling and inspection procedures > FDA: Resin conforms to FDA 21 CFR 177.1520 > CONEG / RoHS: Plastics and colorants are in conformity with CONEG / RoHS standards for heavy metals

> REACH (SVHC): Plastic is in conformity to REACH standards

> LATEX: Material is Latex Free **▶** BSE / TSE: Material is BSE / TSE Free

5. **Quality Assurance:**

- Clear, no presence of contamination in plastic
- Visual attributes
- Closure verification

Document Template: FRM-106 Rev00 Technical Data Sheet

Issue Date: 2016-04-26

Edition Date: 2014-05-21 **Revision Date:** 2020-08-04

Revised by: Annette Roy, QA Coordinator

6. Traceability:

Lot No. Composition: 8 to 9 digits

➤ The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

7. Storage Conditions:

- Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

8. Recommended Use:

- Verify proper cap closure when using biohazard material and / or chemical reagents;
- > Follow chemical resistance chart recommendations;

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev00 Technical Data Sheet

Issue Date: 2016-04-26