

Edition Date: Revision Date : Revised by:

Product Name: Micrewtube[®] Catalogue No: T332-6S

This document replaces any previous version

- **1. Product Description:**
 - 2ml Micrewtube[®]: Sterile, self-standing tube with Silicone washer seal screw cap and attachment loop. Caps are pre-attached and screwed on.
- 2. Packaging:
 - Case: 10 bags of 50 for a total of 500 units per case
- 3. Product Specifications:
 - > Material:
 - Polypropylene tube
 - Polypropylene cap
 - Silicone washer seal
 - Certified RNase, DNase, Pyrogen and DNA Free
 - ➤ Gamma radiation sterilized at a SAL of 10⁻³; specified dose between 6.5 kGy and 13.5 kGy
 - Temperature range: -196°C to +121°C.
 - Autoclavable at 121°C, for up to 30 minutes
 - Centrifuged up to 17 000g

4. Standards and Conformity:

- > ISO 2859-1:
- > FDA:
- > USP:

Resin conforms to USP Class VI Plastics and colorants are in conformity with

Sampling and inspection procedures

Resin conforms to FDA 21 CFR 177.1520

- CONEG: Plastics and colorants are in conformity CONEG standards for heavy metals
- > **REACH (SVHC):** Plastic is in conformity to REACH standards
- LATEX: Material is Latex Free
- **BSE / TSE:** Material is BSE / TSE Free

5. Quality Assurance:

- Clear, no presence of contamination in plastic
- Visual attributes
- Volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg
- Gas phase of Liquid Nitrogen resistance



Edition Date: Revision Date : Revised by: 2005-10-08 2020-05-06 Annette Roy, Regulatory

6. Traceability:

- **Lot No. Composition:** 8 or 9 digits
- > The lot number can be found in one or all of these locations:
 - 1. On exterior case label
 - 2. On label inserted inside the master case
 - 3. On the inner bag

7. Storage Conditions:

- Store at room temperature in normal warehouse conditions
- > Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions
- Should be used only in the gas phase of Liquid Nitrogen

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.