

Edition Date: Revision Date : Revised by: 2004-02-05 2017-11-03 Annette Roy, QA Coordinator

Product Name: Micrewtube[®] cap Catalogu

Catalogue No.: T340_OSL

This document replaces any previous version

1. Product Description:

Micrewtube[®] Screw Cap:

Disposable, non-sterile cap with loop and Silicone washer seal for Micrewtube[®]. Tubes sold separately.

> Available colors:

- Natural, Blue, Green, Lilac, Orange, Red, Yellow, White, Brown
- > Compatible with:
 - T341 Micrewtube[®] Series
- 2. Packaging:
 - > Case:

1 000 units per package

3. Product Specifications:

- > Material:
 - Cap: Polypropylene
 - Washer Seal: Silicone
- Gamma sterilization resistant;
- Temperature range: -196°C to +121°C;
- Certified RNase, DNase, Pyrogen and DNA-free;
- Autoclavable at 121°C, for up to 30 minutes.

4. Standards and Conformity:

ISO 2859-1:	Sampling and inspection procedures
➢ FDA:	Resin conforms to FDA 21 CFR 177.1520
> USP:	Resin conforms to USP Class VI
CONEG / RoHS:	Plastics and colorants are in conformity with
	CONEG / RoHS standards for heavy metals
REACH (SVHC):	Plastic is in conformity to REACH standards



Edition Date: Revision Date : Revised by: 2004-02-05 2017-11-03 Annette Roy, QA Coordinator

5. Quality Assurance:

- > Clear, no presence of contamination in plastic
- Visual attributes
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg
- Gas phase of Liquid Nitrogen resistance

6. Traceability:

Lot No. Composition: 8 or 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- > Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions
- > Should be used only in the gas phase of Liquid Nitrogen

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.