

# Product Name: Micrewtube<sup>®</sup> C

# Catalogue No.: T341-2T

### This document replaces any previous version

### **1.** Product Description:

- O.5 ml Plain Micrewtube<sup>®</sup>: Non-sterile, self-standing, disposable tube.
  Cap sold separately
- > Available colors:
  - Natural and Brown
- > Compatible with:
  - T340 Cap Series

# 2. Packaging:

Case: 1 000 units per package

### **3. Product Specifications:**

- Material: Polypropylene
- Certified RNase, DNase, Pyrogen and DNA Free
- > Temperature range: -196°C to +121°C. Autoclavable at 121°C, for up to 30 minutes
- Tubes have plain configuration
- Centrifuged up 17 000 g

# 4. Standards and Conformity:

$\triangleright$	ISO 2859-1:	Sampling and inspection procedures
$\triangleright$	FDA:	Resin conforms to FDA 21 CFR 177.1520
$\triangleright$	USP:	Resin conforms to USP Class VI
$\triangleright$	CONEG / RoHS:	Plastics and colorants are in conformity with
		CONEG / RoHS standards for heavy metals
$\triangleright$	REACH (SVHC):	Plastic is in conformity to REACH standards
$\triangleright$	LATEX:	Material is Latex Free
$\triangleright$	BSE / TSE:	Material is BSE / TSE Free

# 5. Quality Assurance:

- Clear, no presence of contamination in plastic
- Visual attributes
- Volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg
- Gas phase of Liquid Nitrogen resistance



**Port** Technical Data Sheet

Edition Date: Revision Date: Revised by:

2013-05-28 2020-12-16 Annette Roy, QA Coordinator

# 6. Traceability:

**Lot No. Composition:** 8 to 9 digits

### > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

#### 7. Storage Conditions:

- Store at room temperature in normal warehouse conditions
- Protect from any possible contamination
- Protect from any damage to the packaging

#### 8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions
- > Should be used only in the gas phase of Liquid Nitrogen

#### **OWNERSHIP OF MATERIALS**:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.