**Edition Date:** 2017-02-21 **Revision Date:** 2017-09-01

Revised by: Annette Roy, QA Coordinator

Product Name: Micrewtube® Catalogue No.: T361-6T

This document replaces any previous version

#### 1. **Product Description:**

> 2ml Micrewtube® with Molded Ridges: Disposable, non-sterile, plain, self-standing tube . Caps

sold separately.

> Compatible with:

• T340 Micrewtube® Cap Series

2. Packaging:

> Case: 1 000 units per package

## **Product Specifications:** 3.

Material:

• Tube: Polypropylene

- Certified RNase, DNase, Pyrogen and DNA Free
- > Temperature range: -196°C to +121°C
- ➤ Autoclavable at +121°C, for up to 30 minutes
- > These tubes have molded ridges matching the hole configurations of *Simport* T360 OneHand<sup>™</sup> Rack
- > Centrifuge resistant at up to 17 000 g

### 4. **Standards and Conformity:**

> ISO 2859-1: Sampling and inspection procedures > FDA: Resin conforms to FDA 21 CFR 177.1520

➤ USP: Resin conforms to USP Class VI

> CONEG / RoHS: Plastics and colorants are in conformity with

CONEG / RoHS standards for heavy metals

> REACH (SVHC): Plastic is in conformity to REACH standards

> LATEX: Material is Latex Free ➤ BSE / TSE: Material is BSE / TSE Free

Document Template: FRM-106 Rev01 Technical Data Sheet Page 1 / 2 Issue Date: 2017-06-15

**Edition Date:** 2017-02-21 **Revision Date:** 2017-09-01

Revised by: Annette Roy, QA Coordinator

#### 5. **Quality Assurance:**

- > Clear, no presence of contamination in plastic
- Visual attributes
- > Volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg
- Gas phase of Liquid Nitrogen resistance

## 6. Traceability:

**Lot No. Composition:** 8 or 9 digits

# > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

### 7. **Storage Conditions:**

- > Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- > Protect from any possible contamination
- Protect from any damage to the packaging

#### 8. **Recommended Use:**

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions
- > Should be used only in the gas phase of Liquid Nitrogen

## **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Page 2 / 2 Issue Date: 2017-06-15