**Edition Date:** 2006-03-28 Revision Date: 2020-08-26

Revised by: Annette Roy, Regulatory

Product Name: Cultubes™ Catalogue No.: T405-2

This document replaces any previous version

#### 1. **Product Description:**

> 5ml Culture Tube with Cap: Disposable, sterile 12 x 75mm culture tube with graduations and

white marking area

2. Packaging:

> Case: 20 trays of 25 units / 500 units per case

### 3. **Product Specifications:**

Material:

• Tube: Polystyrene • Cap: Polyethylene > Dual-position snap cap:

- Sealed position for anaerobic cultures
- Loose position for aerobic work
- Certified Pyrogen Free
- ➤ Gamma radiation sterilized at a SAL of 10<sup>-3</sup>; specified dose between 6.5 kGy and 13.5 kGy
- > Temperature range for tube only: -80°C to +80°C
- Not autoclavable
- > Centrifuge resistant up to 1 400 g

### **Standards and Conformity:** 4.

> ISO 2859-1: Sampling and inspection procedures

> FDA: Resins conform to FDA 21 CFR 177.1640 & CFR 177.1520

> CONEG / RoHS: Plastics and colorants are in conformity with CONEG / RoHS standards for heavy metals

> REACH (SVHC): Plastics are in conformity to REACH standards

> LATEX: Materials are Latex Free

### 5. **Quality Assurance:**

- > Clear, no presence of contamination in plastic
- Visual attributes
- > Dimension and volume measurements
- Closure verification

Page 1 / 2 Issue Date: 2017-06-15

Document Template: FRM-106 Rev01 Technical Data Sheet

**Edition Date:** 2006-03-28 **Revision Date:** 2020-08-26

**Revised by:** Annette Roy, Regulatory

## 6. Traceability:

**Lot No. Composition:** 8 or 9 digits

### > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

# 7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- > Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging which could compromise the product sterility

### 8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions

## **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet
Page 2 / 2

Issue Date: 2017-06-15