

## Technical Data Sheet

**Edition Date:** 2017-08-31 **Revision Date:** 2020-08-26

**Revised by:** Annette Roy, Regulatory

Product Name: Cultubes<sup>™</sup> Catalogue No.: T406-33

This document replaces any previous version

## 1. Product Description:

14ml Culture Tubes with Cap: Disposable, sterile 17 x 95mm culture tube, bulk with

graduations and white marking area with cap

2. Packaging:

Case: 500 units per case

### 3. Product Specifications:

- Material:
  - Tube; Polystyrene;
  - Cap: Polyethylene.
- Two-position snap cap:
  - Sealed position for anaerobic cultures;
  - Loose position for aerobic work.
- Certified Pyrogen Fee;
- ➤ Gamma radiation sterilized at a SAL of 10<sup>-3</sup>; specified dose between 6.5 kGy and 13.5 kGy;
- > Temperature range: -80°C to +80°C;
- Not autoclavable;
- > Tube with marking area and printed graduations;
- Centrifuge resistant up to 1 400 g.

#### 4. Standards and Conformity:

> ISO 2859-1: Sampling and inspection procedures

FDA: Resin conforms to FDA 21 CFR 177.1640 & CFR 177.1520

➤ **CONEG:** Plastics and colorants are in conformity with

CONEG standards for heavy metals

REACH (SVHC): Plastic is in conformity to REACH standards

➤ LATEX: Material is Latex Free

## 5. Quality Assurance:

- Clear, no presence of contamination in plastic;
- Visual attributes;
- Dimension and volume measurements;
- Closure verification.

Document Template: FRM-106 Rev00 Technical Data Sheet

Issue Date: 2016-04-26



# Scientific Technical Data Sheet

**Edition Date:** 2017-08-31 **Revision Date:** 2020-08-26

**Revised by:** Annette Roy, Regulatory

## 6. Traceability:

**Lot No. Composition:** 8 to 9 digits

#### > The lot number can be found in one or all of these locations:

- 1. On exterior case label;
- 2. On label inserted inside the master case;
- 3. On the inner bag.

## 7. Storage Conditions:

- Store at room temperature in normal warehouse conditions;
- Avoid temperature variations and humidity;
- Protect from any possible contamination;
- Protect from any damage to the packaging, which could compromise the product sterility.

#### 8. Recommended Use:

- Verify proper cap closure when using biohazard material and / or chemical reagents;
- > Follow chemical resistance chart recommendations;
- > For use in automated equipment, follow the equipment manufacturer's instructions.

#### **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev00 Technical Data Sheet

Issue Date: 2016-04-26