

Scientific Inc. Technical Data Sheet

Edition Date: 2016-11-03 Revision Date: 2022-04-26

Revised by: Christine Laforge RT, Technical Representative

Product Name: Tamper Evident

Transport Tube

Catalogue No.: T552-10ATTP

This document replaces any previous version

1. **Product Description:**

> 10ml Tamper Evident Transport Tube: Disposable, non-sterile, self-standing and non-graduated.

Caps sold separately.

> Compatible with:

• T552WTP Tamper Evident Cap with Lip Seal

• T552W Lip Seal Cap

2. Packaging:

Case: 1 000 units per pack

3. **Product Specifications:**

> Material: Polypropylene

> Temperature range: -90°C to +121°C

> Autoclavable at 121°C, for up to 30 minutes

> Tube has a tamper evident collar

Gamma radiation resistant

4. **Standards and Conformity:**

> ISO 2859-1: Sampling and inspection procedures > FDA: Resin conforms to FDA 21 CFR 177.1520 > CONEG / RoHS: Plastics and colorants are in conformity with CONEG / RoHS standards for heavy metals

> REACH (SVHC): Plastic is in conformity to REACH standards

> LATEX: Material is Latex Free ➤ BSE / TSE: Material is BSE / TSE Free

5. **Quality Assurance:**

- Clear, no presence of contamination in plastic
- Visual attributes
- Dimensional and volume measurements
- Closure verification
- Gas phase of Liquid Nitrogen resistance
- Leak proof testing in vacuum at 71.3cm Hg (95 kPa tested)

Page 1 / 2 Issue Date: 2017-06-15

Document Template: FRM-106 Rev01 Technical Data Sheet

Edition Date: 2016-11-03 **Revision Date:** 2022-04-26

Revised by: Christine Laforge RT, Technical Representative

6. Traceability:

Lot No. Composition: 8 or 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

7. **Storage Conditions:**

- > Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- > Protect from any possible contamination
- Protect from any damage to the packaging

8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- For use in automated equipment, follow the equipment manufacturer's instructions

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are Simport Scientific's copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet Page 2 / 2

Issue Date: 2017-06-15