

# Scientific Inc. Technical Data Sheet

**Edition Date:** 2016-11-03 Revision Date: 2022-04-27

Revised by: Christine Laforge RT, Technical Representative

**Product Name: Tamper Evident** 

**Transport Tube** 

Catalogue No.: T552-12ATTP

This document replaces any previous version

#### 1. **Product Description:**

> 12 ml Tamper Evident Transport Tube: Disposable, non-sterile, self-standing and non-graduated.

Caps sold separately.

# > Compatible with:

- T552WTP Tamper Evident Cap with Lip Seal
- T552W Lip Seal Cap

#### 2. Packaging:

Case: 4pks of 250 units/1 000 units per case

#### 3. **Product Specifications:**

> Material: Polypropylene

> Temperature range: -90°C to +121°C

- > Autoclavable at 121°C, for up to 30 minutes
- > Tube has a tamper evident collar
- Gamma radiation resistant

#### 4. **Standards and Conformity:**

> ISO 2859-1: Sampling and inspection procedures > FDA: Resin conforms to FDA 21 CFR 177.1520 > CONEG / RoHS: Plastics and colorants are in conformity with CONEG / RoHS standards for heavy metals

> REACH (SVHC): Plastic is in conformity to REACH standards

> LATEX: Material is Latex Free ➤ BSE / TSE: Material is BSE / TSE Free

#### 5. **Quality Assurance:**

- Clear, no presence of contamination in plastic
- Visual attributes
- Dimensional and volume measurements
- Closure verification
- Gas phase of Liquid Nitrogen resistance
- Leak proof testing in vacuum at 71.3cm Hg (95 kPa tested)

Page 1 / 2 Issue Date: 2017-06-15

Document Template: FRM-106 Rev01 Technical Data Sheet

**Edition Date:** 2016-11-03 **Revision Date:** 2022-04-27

**Revised by:** Christine Laforge RT, Technical Representative

## 6. Traceability:

**Lot No. Composition:** 8 or 9 digits

### > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

# 7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- > Protect from any possible contamination
- Protect from any damage to the packaging

### 8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions

### **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet
Page 2 / 2

Issue Date: 2017-06-15