

# Scientific Inc. Technical Data Sheet

**Edition Date:** 2013-10-01 **Revision Date:** 2022-04-26

**Revised by:** Christine Laforge RT, Technical Representative

Product Name: Tamper Evident Cap Catalogue No.: T552-7WTP

This document replaces any previous version

## 1. Product Description:

➤ Tamper evident Cap with Inner Lip-Seal: Disposable, non-sterile cap for transport tube.

Transport tubes sold separately. Cap configuration allows insertion of a Capinsert (T345 series) for color-coding

- > Compatible with:
  - T552-7ATTP Tamper Evident Transport Tube
- 2. Packaging:

> Case: 1 000 units per package

# 3. Product Specifications:

➤ Material:

• Cap: High Density Polyethylene

Cap Color: White

> Temperature range: -90°C to +100°C

Gamma radiation resistant

### 4. Standards and Conformity:

➤ ISO 2859-1: Sampling and inspection procedures
 ➤ FDA: Resin conforms to FDA 21 CFR 177.1520
 ➤ CONEG / RoHS: Plastics and colorants are in conformity with CONEG / RoHS standards for heavy metals

> LATEX: Material is Latex Free

Document Template: FRM-106 Rev01 Technical Data Sheet

Issue Date: 2017-06-15

**Edition Date:** 2013-10-01 **Revision Date:** 2022-04-26

**Revised by:** Christine Laforge RT, Technical Representative

## 5. Quality Assurance:

- Clear, no presence of contamination in plastic
- Visual attributes
- Dimensional and volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg (95 kPa tested)

## 6. Traceability:

**Lot No. Composition:** 8 or 9 digits

#### > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

# 7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- > Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

#### 8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions

#### **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet
Page 2 / 2

Issue Date: 2017-06-15