

# Scientific inc Technical Data Sheet

**Edition Date:** 2022-04-06 **Revision Date:** None

Edited by: Christine Laforge RT, Technical Representative

Page 1 / 2

Catalogue No.: T341-2TBR Product Name: Micrewtube®

This document replaces any previous version

1. **Product Description:** 

> 0.5 ml Micrewtube®: Non-sterile, Disposable Self-Standing tube. Tinted

for light sensitive samples. Caps sold separately

> Compatible with:

• T340 Cap Series

2. Packaging:

> Case: 1 000 units per bag

**Product Specifications:** 3.

➤ Material:

Polypropylene

> Color:

• Brown

Certified RNase, DNase, Pyrogen and DNA Free

> Temperature range: -196°C to +121°C

> Autoclavable at 121°C, for up to 30 minutes

Centrifuge resistant at up to 20 000 g

#### 4. **Standards and Conformity:**

> ISO 2859-1: Sampling and inspection procedures > FDA: Resin conforms to FDA 21 CFR 177.1520

> USP: Resin conforms to USP Class VI

> CONEG / RoHS: Plastics and colorants are in conformity with

CONEG / RoHS standards for heavy metals

> REACH (SVHC): Plastic is in conformity to REACH standards

> LATEX: Material is Latex Free ➤ BSE / TSE: Material is BSE / TSE Free

Document Template: FRM-106 Rev01 Technical Data Sheet

Issue Date: 2017-06-15

Edition Date: 2022-04-06 Revision Date: None

Edited by: Christine Laforge RT,
Technical Representative

Page 2 / 2

## 5. Quality Assurance:

- Clear, no presence of contamination in plastic
- Visual attributes
- Volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg (equivalent of 95kPa)
- Gas phase of Liquid Nitrogen resistance

### 6. Traceability:

Lot No. Composition: 8 or 9 digits

#### > The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

# 7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- Protect from any possible contamination
- Protect from any damage to the packaging

#### 8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- For use in automated equipment, follow the equipment manufacturer's instructions
- Should be used only in the gas phase of Liquid Nitrogen

#### **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet

Issue Date: 2017-06-15