

Edition Date: Revision Date : Edited by: 2022-04-06 None Christine Laforge RT, Technical Representative

## Product Name: Micrewtube<sup>®</sup> Catalogue No.: T341-7TBR

### This document replaces any previous version

**1.** Product Description:

# 2.0 ml Micrewtube<sup>®</sup>: Non-sterile, Disposable Conical bottom tube. Tinted for light sensitive samples. Caps sold separately

- > Compatible with:
  - T340 Cap Series

## 2. Packaging:

Case:

1 000 units per bag

## **3. Product Specifications:**

- > Material:
  - Polypropylene
- ➤ Color:
  - Brown
- Certified RNase, DNase, Pyrogen and DNA Free
- Temperature range: -196°C to +121°C
- > Autoclavable at 121°C, for up to 30 minutes
- Centrifuge resistant at up to 20 000 g

## 4. Standards and Conformity:

## ISO 2859-1:Sampling and inspection procedures

FDA: Resin conforms to FDA 21 CFR 177.1520
USP: Resin conforms to USP Class VI
CONEG / RoHS: Plastics and colorants are in conformity with CONEG / RoHS standards for heavy metals
REACH (SVHC): Plastic is in conformity to REACH standards
LATEX: Material is Latex Free
BSE / TSE: Material is BSE / TSE Free



Edition Date: Revision Date : Edited by: 2022-04-06 None Christine Laforge RT, Technical Representative

## 5. Quality Assurance:

- > Clear, no presence of contamination in plastic
- Visual attributes
- Volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg (equivalent of 95kPa)
- ➢ Gas phase of Liquid Nitrogen resistance

## 6. Traceability:

- Lot No. Composition: 8 or 9 digits
- > The lot number can be found in one or all of these locations:
  - 1. On exterior case label
  - 2. On label inserted inside the master case
  - 3. On the inner bag

## 7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- Protect from any possible contamination
- Protect from any damage to the packaging

#### 8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions
- Should be used only in the gas phase of Liquid Nitrogen

#### **OWNERSHIP OF MATERIALS**:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.