

Scientific me. Technical Data Sheet

Edition Date: 2022-05-10 **Revision Date:** 2022-08-08

Revised by: Christine Laforge RT, **Technical Representative**

Product Name: SpecTainer™ II Catalogue No.: C567-60CYS

This document replaces any previous version

1. **Product Description:**

> > 60 ml Urine Container: Disposable, sterile and assembled with conventional leakproof

> > screw cap

2. Packaging:

> Case: 5 packs of 100 units / 500 units per case

3. **Product Specifications:**

Material:

• Container: Polypropylene

• Cap: Polyethylene • Cap Color: cyan

- ➤ Gamma radiation sterilized at a SAL of 10⁻³; specified dose between 6.5 kGy and 13.5 kGy
- > Temperature range: -80°C to +100°C
- > Label to enter patient's name
- > Tamper evident label for sterility assurance
- Cap configuration allows insertion of a Capinsert[™] (T345 Series) for color-coding

4. **Standards and Conformity:**

> ISO 2859-1: Sampling and inspection procedures > FDA: Resins conform to FDA 21 CFR 177.1520 ➤ RoHS: Plastics and colorants are in conformity with

RoHS standards for heavy metals

Plastics are in conformity to REACH standards > REACH (SVHC):

> LATEX: Materials are Latex Free

5. **Quality Assurance:**

- Clear, no presence of contamination in plastic
- Visual attributes
- Dimension and volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg (95kPa tested)

Page 1 / 2 Issue Date: 2017-06-15

Edition Date: 2022-05-10 **Revision Date:** 2022-08-08

Revised by: Christine Laforge RT, **Technical Representative**

6. **Traceability:**

Lot No. Composition: 8 or 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

7. **Storage Conditions:**

- > Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging which could compromise sterility

8. **Recommended Use:**

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are Simport Scientific's copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet Page 2 / 2 Issue Date: 2017-06-15