**Edition Date:** 2016-10-28 **Revision Date:** 2020-06-11

Revised by: Annette Roy, QA Coordinator

Product Name: SnapTwist<sup>™</sup> Cap for Microtube®

Catalogue No.: T343 LSL

Page 1 / 2

This document replaces any previous version

### 1. **Product Description:**

Non sterile, cap with loop for SnapTwist™ Micrewtube®. ➤ SnapTwist<sup>TM</sup> Cap:

### Available colors:

• Natural, Blue, Green, Lilac, Red, Yellow and White

### 2. Packaging:

Case: 1 000 units per package

### **Product Specifications:** 3.

- Material: High Density Polyethylene;
- Certified RNase, DNase, Pyrogen and DNA Free;
- Temperature range: -90°C to +100°C;
- > Autoclavable: No
- ► Insertion of a Capinsert<sup>TM</sup> (T345 series) allows color coded identification of contents.

### 4. **Standards and Conformity:**

> ISO 2859-1: Sampling and inspection procedures > FDA: Resin conforms to FDA 21 CFR 177.1520 > CONEG: Plastics and colorants are in conformity

with CONEG standards for heavy metals

➤ REACH (SVHC): Plastic is in conformity to REACH standards

> LATEX: Material is Latex Free ➤ BSE / TSE: Material is BSE / TSE Free

### **Quality Assurance:** 5.

- Clear, no presence of contamination in plastic;
- Visual attributes;
- Volume measurements;
- Closure verification

Document Template: FRM-106 Rev00 Technical Data Sheet

Issue Date: 2016-04-26

**Edition Date:** 2016-10-28 **Revision Date:** 2020-06-11

**Revised by:** Annette Roy, QA Coordinator

# 6. Traceability:

**Lot No. Composition:** 8 to 9 digits

# > The lot number can be found in one or all of these locations:

- 1. On exterior case label;
- 2. On label inserted inside the master case;
- 3. On the inner bag.

# 7. Storage Conditions:

- Store at room temperature in normal warehouse conditions;
- > Avoid temperature variations and humidity;
- Protect from any possible contamination;
- Protect from any damage to the packaging.

### 8. Recommended Use:

- Verify proper cap closure when using biohazard material and / or chemical reagents;
- > Follow chemical resistance chart recommendations;
- For use in automated equipment, follow the equipment manufacturer's instructions;
- Should be used only in the gas phase of Liquid Nitrogen.

## **OWNERSHIP OF MATERIALS:**

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev00 Technical Data Sheet

Issue Date: 2016-04-26