Edition Date: 2013-02-28 Revision Date: 2022-08-08

Revised by: Annette Roy, QA Coordinator

Product Name: SpecTainer™ Catalogue No.: C566-90AQSECO

This document replaces any previous version

1. **Product Description:**

> Eco-Friendly, 90 ml Urine Container: Disposable, Sterile and assembled with tamper evident leakproof closure.

2. Packaging:

Case: 4 packs of 100 units / 400 units per case

Product Specifications: 3.

- Material:
 - Container: Specially formulated degradable Polypropylene
 - Cap: Specially formulated degradable Polyethylene
 - Cap Color: Aqua
- > Fully Biodegradable in 9 months to 5 years in organic environment.
- ➤ Gamma radiation sterilized at a SAL of 10⁻³; specified dose between 6.5 kGy and 13.5 kGy
- > Temperature range: -90°C to +100°C
- ➤ Label to enter patient's name
- ➤ Cap configuration allows insertion of a CapinsertTM (T345 Series) for color-coding

4. **Standards and Conformity:**

> ISO 2859-1: Sampling and inspection procedures > FDA: Resins conform to FDA 21 CFR 177.1520 Plastics and colorants are in conformity with ➤ RoHS:

RoHS standards for heavy metals

Plastics are in conformity to REACH standards > REACH (SVHC):

> LATEX: Materials are Latex Free

5. **Quality Assurance:**

- > Clear, no presence of contamination in plastic
- Visual attributes
- Dimension and volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg

Page 1 / 2 Issue Date: 2017-06-15

Document Template: FRM-106 Rev01 Technical Data Sheet

Edition Date: 2013-02-28 **Revision Date:** 2022-08-08

Revised by: Annette Roy, QA Coordinator

6. Traceability:

Lot No. Composition: 8 or 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- Protect from any possible contamination
- Protect from any damage to the packaging which could compromise sterility

8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Issue Date: 2017-06-15

Document Template: FRM-106 Rev01 Technical Data Sheet