

Edition Date: Revision Date : Revised by:

Product Name: Sample Tube Catalogue No.: C572-1

This document replaces any previous version

- 1. Product Description:
 - **30 ml Sample Tube with Cap:** Disposable, sterile, self-standing conical bottom and skirted tube with cap.
- 2. Packaging:
 - > Case:

20 packs of 25 units / 500 units per case

Resins conform to FDA 21 CFR 177.1520

3. Product Specifications:

- > Material:
 - Tube: Polypropylene
 - Cap: Polyethylene
- ▶ Gamma radiation sterilized at a SAL of 10⁻³; specified dose between 6.5 kGy and 13.5 kGy
- Temperature range: -90°C to +100°C
- > Tubes have molded-in graduations from 10ml to 30ml with 5ml increments
- > Cap configuration allows insertion of a Capinsert[™] (T345 Series) for color-coding

4. Standards and Conformity:

- **ISO 2859-1:** Sampling and inspection procedures
- > FDA:
- > CONEG / RoHS: Plastics and colorants are in conformity with
- CONEG / RoHS standards for heavy metals
- **REACH (SVHC):** Plastics are in conformity to REACH standards
- LATEX: Materials are Latex Free
- **BSE / TSE:** Materials are BSE / TSE Free

5. Quality Assurance:

- > Clear, no presence of contamination in plastic
- Visual attributes
- Volume measurements
- Closure verification
- Gas phase of Liquid Nitrogen resistance



Edition Date: Revision Date : Revised by: 2004-02-05 2022-08-08 Annette Roy, Regulatory

6. Traceability:

- **Lot No. Composition:** 8 or 9 digits
- > The lot number can be found in one or all of these locations:
 - 1. On exterior case label
 - 2. On label inserted inside the master case
 - 3. On the inner bag

7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- > Protect from any damage to the packaging which could compromise the product sterility

8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > Follow the equipment manufacturer's instructions
- > Should be used only in the gas phase of Liquid Nitrogen

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.